



**KB03017**  
**Proanthocyanidins**  
**Assay Kit**

**96 well plate**  
**100/200/400 tests**

## Table of contents

1.	General information	1
2.	Technical specifications	2
3.	Materials and storage	3
4.	Introduction	4
5.	Assay principle	4
6.	Assay preparation	5
7.	Sample preparation	7
8.	Assay protocol	8
9.	Data analysis	9
10.	Troubleshooting	11
11.	Additional information	13
12.	Related products	13
13.	Warranties and limitation of liability	14

# 1. General information

## PRECAUTIONS

Please read this manual carefully before beginning the assay.

This product is designed for **research use only**. It is not approved for human or animal use or clinical diagnosis. All chemicals should be handled with care and in accordance with laboratory safety practices. It is recommended to use basic Personal Protective Equipment.

Do not use after the expiration date stated on the packaging.

Do not mix or substitute reagents or materials from other kit batches or vendors.

For the **material safety data sheet** (MSDS) please contact us at [info@bioquochem.com](mailto:info@bioquochem.com)

## TECHNICAL RECOMMENDATIONS

Store reagents as indicated in **Materials and storage** section.

Be sure to keep the bottle capped when not in use.

Let the components reach room temperature (RT) before use.

Immediately before use, gently invert and rotate reagent bottles several times to mix the contents thoroughly.

Avoid foaming or bubbles when mixing or reconstituting components.

Avoid cross contamination of samples or reagents by changing pipette tips between sample, standard and reagent additions.

Be sure to use the optimal microplate for the assay. Flat bottom transparent microplates for UV/VIS applications, and black microplates for fluorescence measurements.

## 2. Technical specifications

### Available sizes

100/200/400 tests

### Required sample volume

10 µL/test

### Compatible samples

Food and beverages

### Type of detection

Colorimetric (640 nm)

### 3. Materials and storage

#### MATERIALS SUPPLIED

Item	No. Tests	Units	Storage
Reagent A	100	1	RT
	200	2	
	400	4	
Reagent B	100	1	RT
	200	2	
	400	4	
Reagent C	100	1	-20 °C
	200	2	
	400	4	
Standard	100	1	4 °C
	200	2	
	400	4	
Transparent 96-Well Microplate	100	1	RT
	200	2	
	400	4	

#### MATERIALS NEEDED BUT NOT SUPPLIED

- Double distilled water (ddH<sub>2</sub>O) as Milli-Q Ultrapure Water
- Labware materials (micropipettes, tubes, stirring/mixing equipment)
- Colorimetric microplate reader – equipped with filter for OD 640 nm

#### STORAGE CONDITIONS

On receipt, store kit components as indicated above. Under these conditions, the reagents are stable in the original packaging until the expiration date indicated on the outside of the box. After reconstitution, standard solutions are unstable in the presence of oxygen. Prepare a fresh set of standards for every use.

## 4. Introduction

Proanthocyanidins (PACs), are polyphenols of high molecular weight widely found in plants and daily foods. They are responsible for many bitter and astringent flavors in food products such as wine, chocolate, beer, and cranberries.


PACs also known as condensed tannins, are the most abundant flavonoids in the human diet. Recent attention has been given to these compounds in the nutraceutical field due to their potential health benefits (e.g. mitigating inflammation, oxidative stress, protecting against cardiovascular disease, etc.)

**BQC Proanthocyanidins (PACs) Assay Kit is an easy, fast, and reliable method to detect these phenolic compounds in food and beverages samples.**

## 5. Assay principle

This Proanthocyanidins (PACs) Assay Kit is based on the 4-dimethylaminocinnamaldehyde (DMAC) colorimetric method. The DMAC reagent reacts specifically with PACs to form a green/blue colored compound with a maximum of absorbance at 640 nm. DMAC reagent does not react with ascorbic acid, hydroxyl-phenylalkyl acids or other flavonoids.

PACs concentration of samples is determined based on a catechin hydrate standard curve (catechin standard provided with the kit) and the results are expressed as Catechin Equivalents (CE,  $\mu\text{g}$  Catechin Hydrate/mL).



*Principle of PACs (DMAC Method) Assay Kit*

## 6. Assay preparation

### REAGENT PREPARATION

All assay reagents not listed below are ready to use as supplied. Allow the reagents to reach room temperature before use.

**DMAC Working Solution:** Add 1 mL of Reagent B in each vial of Reagent C and mix well.

**⚠ CAUTION:** DMAC Solution must be prepared immediately before use.

**Standard Solution (Catechin Hydrate):** Add 1 mL of Reagent A to the Standard vial and mix well. Use this solution to prepare the standard curve.

### STANDARD CALIBRATION

Prepare Catechin Hydrate standards for the calibration curve from the Standard solution according to the following Table. Prepare the standards immediately prior to each assay. Vortex tubes thoroughly. Discard standard solutions after use.

Standard	Standard solution (µL)	Reagent A (µL)	* CE (µg Catechin Hydrate/mL)
Std 1 (Reagent Blank)	0	100	0
Std 2	2.5	197.5	25
Std 3	5	195	50
Std 4	10	190	100
Std 5	15	185	150
Std 6	20	180	200

\* Proanthocyanidins content is expressed as CE (Catechin Equivalents)

## PLATE SET UP

**BQC recommends running the standards and samples at least in duplicate (triplicate recommended).** There is no specific pattern for using the wells on the plate. A proposed layout of standards (Std) and samples (S) to be measured in duplicate is shown below.

**NOTE:** If sample blanks are included in the assay, it is necessary to reserve some wells of the plate for these blanks

	1	2	3	4	5	6	7	8	9	10	11	12
A	Std 1	Std 1	S3	S3	S11	S11	S19	S19	S27	S27	S35	S35
B	Std 2	Std 2	S4	S4	S12	S12	S20	S20	S28	S28	S36	S36
C	Std 3	Std 3	S5	S5	S13	S13	S21	S21	S29	S29	S37	S37
D	Std 4	Std 4	S6	S6	S14	S14	S22	S22	S30	S30	S38	S38
E	Std 5	Std 5	S7	S7	S15	S15	S23	S23	S31	S31	S39	S39
F	Std 6	Std 6	S8	S8	S16	S16	S24	S24	S32	S32	S40	S40
G	S1	S1	S9	S9	S17	S17	S25	S25	S33	S33	S41	S41
H	S2	S2	S10	S10	S18	S18	S26	S26	S34	S34	S42	S42

Example of plate layout for the Proanthocyanidins Assay Kit



## 7. Sample preparation

**The following sample preparation protocols are intended as a guide only. The optimal conditions for sample preparation must be determined by the end user.** It is recommended to use fresh samples. If it is not possible, aliquot and store samples appropriately with minimal freeze/thawing.

**Proanthocyanidins Assay Kit can be used to determine PACs in food, beverages and plant extracts.**

**Food and beverages.** Fruit juices and other beverages such as wine, tea, and coffee can be directly measured with appropriate dilutions. If it is required, clarify the sample through filtration prior performing the assay. Ensure that the selected filter is appropriate for filtering your samples, avoiding polyphenols retention.







For the analysis of other samples like **fruits, vegetables, and plants** an extraction step is usually required. The extraction method varies based upon the sample type. The most common extraction solvents include acid/methanol, acid/ethanol, or acetone.

Reagents and materials required for sample preparation are not supplied with the kit. Before doing sample preparation, consider the volume of sample required per test; see **Technical specifications** section.

Make sure that interfering substances present in the sample do not give a significant background. Run proper blanks as necessary (e.g. sample blank should be always evaluated when working with highly colored samples). It is recommended to assay different sample dilutions to ensure the values fall within the standard curve.

## 8. Assay protocol

Prepare and mix all reagents thoroughly before use. Each standard, sample or blank should be assayed at least in duplicate.


-  Set up the plate design
-  Add **10 µL** of **standard** or **sample** in each well
-  Add **230 µL** of **Reagent A** in all wells
-  Add **10 µL** of **DMAC Solution** in all wells
-  Incubate for **15 minutes** protected from light at RT
-  Read the **absorbance** of all wells at **640 nm** in end point mode at **RT**

If you need to **adapt this kit** for another form of the assay (for example cuvette), **contact us at [info@bioquochem.com](mailto:info@bioquochem.com)**

## 9. Data analysis

### ANALYSIS OF THE STANDARDS

- Calculate the average absorbance of the standards.
- Subtract the average absorbance of the reagent blank (Std 1) from the average absorbance of the standards to obtain the blank-corrected absorbance of the standards.
- Create a standard curve by plotting the blank-corrected absorbance of the standards as a function of the standard concentration (see **STANDARD CALIBRATION** section). A typical standard curve ( $y = \text{slope} \cdot x \pm \text{intercept}$ ) for this assay is shown below.



*CE Standard curve for PACs (DMAC Method) Assay Kit*

- ⚠ This standard curve is an example of the data typically obtained with this kit. **DO NOT USE** this standard curve to calculate the phenolic content of your samples. A new standard curve must be performed by the end user.

## ANALYSIS OF THE SAMPLES

- Calculate the average absorbance of the samples.
- Subtract the average absorbance of the reagent blank (Std 1) from the average absorbance of each sample to obtain the blank-corrected absorbance of the samples ( $A_s$ ).
- Calculate the PACs concentration of the samples expressed as Catechin Equivalents (CE,  $\mu\text{g}$  Catechin Hydrate/mL) using the following equation. Slope and intercept values are obtained from the standard curve.

$$\text{CE } (\mu\text{g Catechin Hydrate/mL}) = \left( \frac{A_s - \text{intercept}}{\text{slope}} \right)$$

When working with diluted samples the CE values obtained must be multiplied by the dilution factor to obtain the PAC concentration of the undiluted sample.

## 10. Troubleshooting

This troubleshooting table provides potential sources and solutions for common problems observed with BQC Assay Kits. **The problems listed below could occur when using any BQC Assay Kit.** They are not specific for this assay kit.

Problem	Possible Cause	Recommended Solution
Wells have color but there is no reading	Plate read at incorrect wavelength	Check the wavelength used in the assay
	Incorrect microplate	Use the correct microplate for your application UV/Vis: transparent Fluorescence: black wells/transparent bottom
Standard readings do not follow a linear pattern	Pipetting errors in preparation of standards	Avoid pipetting small volumes (<5 $\mu$ L) Be careful not to splash from well to well
	Air bubbles formed in well(s)	Use reverse pipetting technique
	Standard stock is at incorrect concentration	Always refer to dilutions described in <b>Assay preparation</b>
	Improperly thawed reagents	Thaw all components completely and mix well before use
	Use of improperly stored reagents	Store the components appropriately Use fresh components from the standard curve
	Incorrect incubation times or temperatures	Refer to <b>Assay protocol</b>
Dispersion of standard and sample readings	Pipetting errors	Avoid pipetting small volumes (<5 $\mu$ L) Be careful not to splash from well to well
	Air bubbles formed in well(s)	Use reverse pipetting technique

Problem	Possible Cause	Recommended Solution
Sample erratic values	Samples contain interfering substances	Dilute sample further (if possible)
	Inappropriately stored samples or samples used after multiple freeze-thaw cycles	Use fresh samples or store appropriately until use
	Samples not deproteinized	Use an appropriate deproteinization protocol
	Cells/Tissue samples not homogenized completely	Repeat the sample homogenization
	Inappropriate sample dilution buffer	Refer to <b>Assay preparation</b>
Sample reading fall outside the detection range	Samples are too diluted/concentrated No analyte/activity is observed in the sample	Re-assay using different sample dilutions

## STILL HAVING PROBLEMS?

Contact BQC if you have any further questions, our team will be pleased to help you:



Phone

+ 34 985 26 92 92



E-mail

[info@bioquochem.com](mailto:info@bioquochem.com)


Business hours

 Monday-Thursday: 8.30 to 17.00 (CEST)  
 Friday: 8.00 to 15.00 (CEST)

## 11. Additional information

**Proanthocyanidins Assay Kit** is a quick (< 30 minutes) and precise (RSD < 8 %) assay for determining PACs in a wide variety of samples.

If unexpected results are obtained running your samples, please contact us at [info@bioquochem.com](mailto:info@bioquochem.com)

## 12. Related products

More products available on [bioquochem.com](http://bioquochem.com)

Reference	Product
KB03006	Polyphenols Quantification Assay Kit
KF01001	DMPD Antioxidant Capacity Assay Kit
KB03007	Thiol Quantification Assay Kit

## 13. Warranties and limitation of liability

BQC shall not in any event be liable for incidental, consequential or special damages of any kind resulting from any use or failure of the products, even if BQC has been advised of the possibility of such damage including, without limitation, liability for loss of use, loss of work in progress, downtime, loss of revenue or profits, failure to realize savings, loss of products of buyer or other use or any liability of buyer to a third party on account of such loss, or for any labor or any other expense, damage or loss occasioned by such product including personal injury or property damage is caused by BQC's gross negligence. Any and all liability of BQC hereunder shall be limited to the amounts paid by the buyer for the product.

Buyer's exclusive remedy and BQC's sole liability hereunder shall be limited to a refund of the purchase price, or the replacement of all material that does not meet our specifications.

Said refund or replacement is conditioned on buyer giving written notice to BQC within 30 days of shipment.

**Expiration date:** 1 year from the date of fabrication. Expiration date is indicated on the outside of the box.

For further details, please refer to our website [bioquochem.com](http://bioquochem.com)





Vivero Ciencias de la Salud  
C. Colegio Santo Domingo de Guzmán  
33011 Oviedo, Asturias, Spain



[www.bioquochem.com](http://www.bioquochem.com)